

Minutes of Pre Bid Meeting

Date:-28/8/2007.

Pre Bid meeting for Tender No. 35,36,37,41 & 42 was held on 27/8/2007 at 11:00am in the conference hall of DTE. The following members were present.

1. Prof G.B Dhanokar	Joint Director, DTE
2. Shri A. V. Amte	ITE, DTE
3. Prof Salunke	G. P. Thane
4. Shri G. Chandak	Xtreme solution, Pune
5. Shri Kedar Apte	Director, Eternal Engg Equipments . Pune
6. Shri Sachin Sontake	New Anukool Kolhapur
7. Shri Abhijeet Kulkarni	New Anukool Kolhapur

The following points were discussed during the meeting:

1. Tender No. 35. It was pointed out that word 'Eddy' should be used in place of Body before current dynamometer. The printing mistake was accepted. Also it was accepted that in place of 'orifice', the word 'arifice' was printed, which may be a printing mistake. This mistake was accepted.
2. Is dynamometer in tender 35 required of a specific brand ?
No any suitable brand which meets the requirements of system will be acceptable
3. It was informed that engine should be brand new. Refurbished engine should not be used.
4. Tender No. 36 compressor should be ½ tonne hermetically sealed, reciprocating type. Refrigerant should be 134-a
Pressure sensors should be used in place of Pressure gauges in point 10
5. Word 'COP' should be read as co-efficient of performance in all related tender specifications.
6. Tender 41:- Compressor should be ½ tonne in place of 1 tonne.
VCRS means 'Vapour Compression Refrigeration system'.
7. In tender no.41, refrigerant should be R134-a in place of 'CFC free Refrigerant'. The following two lines should be added to the tender specification:
'should be able to plot pressure enthalpy and temperature entropy plot'
'should be able to calculate actual co-efficient of performance [COP] and the theoretical co-efficient of performance.'
8. In Tender No 42, lithium bromide is accepted in place of 'ammonia & water'.
Also the capacity should be 1 tonne in place of ITR.
Word Eureka should be deleted from 'Rotameter Eureka make'
Delete the words 'sight glass for liquid ammonia' since ammonia based system is not required.
Instead of MSBS, read MCBs. i.e. miniature circuit breakers
Word displaying is required in place of depicting
Word schematic is required in place of scheme.
9. In tender No 39 word computerized required in place of computer controlled
10. In Tender No 44, delete word electric from electric dynamometer.
